



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900

Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Reid Reagan
Browning High School - Lead in Schools
105 US Hwy 89
Browning, MT 59417

PWS ID: MTOPI0543
Project: Lead in Schools

Client Sample ID: FX002 **Lab ID:** 2203721-01
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX004 **Lab ID:** 2203721-02
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX008 **Lab ID:** 2203721-03
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX009 **Lab ID:** 2203721-04
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX013 **Lab ID:** 2203721-05
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Reid Reagan
Browning High School - Lead in Schools
105 US Hwy 89
Browning, MT 59417

PWS ID: MTOPI0543
Project: Lead in Schools

Client Sample ID: FX016 **Lab ID:** 2203721-06
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX021 **Lab ID:** 2203721-07
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX024 **Lab ID:** 2203721-08
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX025 **Lab ID:** 2203721-09
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX028 **Lab ID:** 2203721-10
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900

Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Reid Reagan
Browning High School - Lead in Schools
105 US Hwy 89
Browning, MT 59417

PWS ID: MTOPI0543
Project: Lead in Schools

Client Sample ID: FX043 **Lab ID:** 2203721-11
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX048 **Lab ID:** 2203721-12
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW

Client Sample ID: FX060 **Lab ID:** 2203721-13
Matrix: DRINKING WATER **Collected:** 04/13/2022 9:02 **Received:** 04/19/2022 11:15

<u>Analyses</u>	<u>Result</u>	<u>Units</u>	<u>MDL</u>	<u>MCL</u>	<u>Method</u>	<u>Prepared</u>	<u>Analyzed</u>	<u>Analyst</u>
Lead in schools	ND	mg/L	0.001		E200.8		05/06/2022 9:25	BLW



ANALYTICAL REPORT

Montana Environmental Laboratory LLC

1170 N. Meridian Rd., P.O. Box 8900, Kalispell, MT 59904-1900
 Phone: 406-755-2131 Fax: 406-257-5359 www.melab.us

Reid Reagan
 Browning High School - Lead in Schools
 105 US Hwy 89
 Browning, MT 59417

PWS ID: MTOPI0543
 Project: Lead in Schools

3721 pg.1

3721

MONTANA ENVIRONMENTAL LAB, LLC
 P.O. BOX 8900, KALISPELL, MT 59904
 VOICE: (406) 755-2131 • FAX: (406) 257-5359
 www.melab.us • info@melab.us

Chain Of Custody and Analytical Request Record

Please Print. Provide as much information as possible

Page 1 of 2

Company Name: <u>Dist#9</u> <u>Browning Public Schools / Browning High School</u> Report Address: <u>105 US Hwy 89</u> <u>Browning MT 59417</u> MH OPI-0543	Project Name, PWSID #, Permit #, Etc. <u>Lead in Schools</u> Contact Name, Phone, Fax: <u>Reid Reagan (406) 338-2745</u> Invoice Contact & Phone #: <u>FAX 338-5216</u>
Invoice Address: _____ Purchase Order # _____	TURN AROUND TIME REQUIRED <input type="checkbox"/> STANDARD 10 WORKING DAYS <input type="checkbox"/> 5 WORKING DAYS (50% SURCHARGE) <input type="checkbox"/> 1-2 WORKING DAYS (100% SURCHARGE) <input type="checkbox"/> OTHER _____

Alvin Vellon Owl III *Alvin Vellon Owl III*
 SAMPLER PRINTED NAME AND SIGNATURE:

	Analysis Requested										COMMENTS
	1	2	3	4	5	6	7	8	9	10	
Lead In Schools											INV DEQ
Browning High School											TESTS TAKEN DURING BOS
Browning H.S											Spring Break mid
Browning H.S											Morning But with
Browning H.S											Schools / Kitchens
Browning H.S											Empty would be first
Browning H.S											flow of the morning or
Browning H.S											DAY
Browning H.S											
Browning H.S											
Browning H.S											

FX#	LAB#	SITE NAME or LOCATION	RESULTS	SAMPLE DATE	SAMPLE TIME	MATRIX*	# Bottles
				4-13-22	9:02AM		X
				4-13-22	9:05AM		X
				4-13-22	9:15AM		X
				4-13-22	9:18AM		X
				4-13-22	9:23AM		X
				4-13-22	9:27AM		X
				4-13-22	9:35AM		X
				4-13-22	9:48AM		X
				4-13-22	9:58AM		X
				4-13-22	10:15AM		X

Custody Record MUST be Signed	Relinquished by: <i>Alvin Vellon Owl III</i>	Date/Time: <i>4/19/22 11:21 AM</i>	Shipped by:	Received by: <i>JB</i>	Date/Time: <i>4-19-22 11:15</i>
	Relinquished by:	Date/Time:	Shipped by:	Received by:	Date/Time:
	Relinquished by:	Date/Time:	Shipped by:	Received by:	Date/Time:

NET 30 DAYS: 1.5% PER MONTH INTEREST CHARGE (18% A.P.R.) CUSTOMER AGREES TO PAY COLLECTION COSTS AND ATTORNEY'S FEES.

- INSTRUCTIONS**
1. USE ONE LINE PER SAMPLE.
 2. BE SPECIFIC IN TEST REQUESTS
 3. CHECK OFF TESTS TO BE PERFORMED FOR EACH SAMPLE

LAB USE ONLY:

LOGIN COMMENTS: _____	SAMPLES CHECKED/LOGGED IN BY: _____ SAMPLE TEMP, RECEIPT AT LAB: _____ (Compliance: 4 +/- 2°C) SAMPLES RECEIVED DAY OF COLLECTION? <input type="checkbox"/> (check for yes) BLUE ICE: FROZEN PARTIALLY FROZEN THAWED
-----------------------	---

